

TEST ThAr FRAMES AFTER THE HCFA MAINTENACE
June 26, 2006

With spring in between the translation stages. Vacuum vessel pump off.

Tel. Pos.	Drift A m/s	Drift B m/s	Flux ratio A	Flux ratio B
Z	0.240	0.404	1.000	1.000
Z	0.053	-0.026	1.000	1.000
N60	-0.043	-0.058	0.998	0.994
N60	-0.247	-0.002	1.000	1.000
Z	0.096	0.113	1.001	1.005
Z	0.232	0.054	1.001	1.000
S60	-0.032	-0.147	0.980	0.999
S60	0.063	0.150	0.999	1.000
Z	0.018	-0.100	1.013	1.000
Z	-0.023	-0.026	1.000	1.000
E60	-0.330	0.066	1.003	1.002
E60	0.196	0.046	1.000	1.000
Z	0.463	0.174	1.003	0.997
Z	0.391	0.167	1.003	0.998
W60	0.399	0.013	1.022	1.004
W60	-0.183	-0.302	1.000	1.001
Z	-0.473	0.083	0.983	0.997
Z	0.167	0.126	1.000	0.999